



Corporate Accreditation No 63  
Accredited for compliance with ISO/IEC 17025 - Testing



## Analytical Report 287892

Issue Date: 14/07/2023  
Issued By : Sydney Water Laboratory Services

Delivery Address: Sydney Water Corporation  
[Redacted]  
West Ryde NSW 2114

Telephone: [Redacted]  
Email: [Redacted]

Attention: [Redacted]  
Customer: Department of Planning and Environment  
Customer ID: ZOEH\_1

Address: [Redacted]  
Telephone: [Redacted]  
Email: [Redacted]

### CONTENTS

- 1. Sydney Water Approved Signatory
- 2. Sample Summary
- 3. Analytical results
- 4. Comments
- 5. Laboratory QC results

### Sydney Water Approved Signatory

[Redacted], Phycology Supervisor	[Redacted], Phycology Analyst	[Redacted], Organics Analyst
----------------------------------	-------------------------------	------------------------------

Where a result is required to meet a compliance limit or specification the associated uncertainty must be considered. Uncertainty estimates are available for all accredited test results.

**SAMPLE SUMMARY**

<u>Client Sample ID</u>	<u>Sample Number</u>	<u>Sampling Procedure</u>	<u>Date Sampled</u>	<u>Date Received</u>	<u>Date Authorised</u>	<u>Description</u>
235784	L23053891	1	27/06/2023	29/06/2023	10/07/2023	B1 (ENVIRONMENTAL WATER)
235785	L23053892	1	27/06/2023	29/06/2023	10/07/2023	B2 (ENVIRONMENTAL WATER)
235786	L23053893	1	27/06/2023	29/06/2023	10/07/2023	B3 (ENVIRONMENTAL WATER)
235787	L23053894	1	27/06/2023	29/06/2023	10/07/2023	E1 (ENVIRONMENTAL WATER)
235788	L23053895	1	27/06/2023	29/06/2023	10/07/2023	E2 (ENVIRONMENTAL WATER)
235789	L23053896	1	27/06/2023	29/06/2023	10/07/2023	E3 (ENVIRONMENTAL WATER)
235790	L23053897	1	27/06/2023	29/06/2023	10/07/2023	E4 (ENVIRONMENTAL WATER)
235791	L23053898	1	27/06/2023	29/06/2023	10/07/2023	E5 (ENVIRONMENTAL WATER)
235800	L23053899	1	27/06/2023	29/06/2023	13/07/2023	B1 (ENVIRONMENTAL WATER)
235801	L23053900	1	27/06/2023	29/06/2023	13/07/2023	B2 (ENVIRONMENTAL WATER)
235802	L23053901	1	27/06/2023	29/06/2023	13/07/2023	B3 (ENVIRONMENTAL WATER)
235803	L23053902	1	27/06/2023	29/06/2023	13/07/2023	E1 (ENVIRONMENTAL WATER)
235804	L23053903	1	27/06/2023	29/06/2023	13/07/2023	E2 (ENVIRONMENTAL WATER)
235805	L23053904	1	27/06/2023	29/06/2023	13/07/2023	E3 (ENVIRONMENTAL WATER)
235806	L23053905	1	27/06/2023	29/06/2023	13/07/2023	E4 (ENVIRONMENTAL WATER)
235807	L23053906	1	27/06/2023	29/06/2023	13/07/2023	E5 (ENVIRONMENTAL WATER)

**Sampling procedures**

- 1 Samples analysed as received.
- 2 Samples collected as per FS procedures SAWI 070, Excluding Oil & Grease which is collected as per clients instructions.
- 3 Samples collected as per FS procedures SAWI 070.
- 4 Results reported as received from WNSW.

**ANALYTICAL RESULTS**

Client Sample ID		235784	235785	235786	235787	235788	235789	235790	235791
Sampled Date		27/06/2023 03:07:00 PM	27/06/2023 11:42:00 AM	27/06/2023 02:00:00 PM	27/06/2023 11:10:00 AM	27/06/2023 10:30:00 AM	27/06/2023 09:55:00 AM	27/06/2023 12:15:00 PM	27/06/2023 12:53:00 PM
Sample Number		L23053891	L23053892	L23053893	L23053894	L23053895	L23053896	L23053897	L23053898
<b>ORGANICS</b>									
TC0049DW : Algal Toxins									
Cylindrospermopsin (extra cellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Cylindrospermopsin (intra cellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Anatoxin-a(extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Anatoxin-a(intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Microcystin RR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05

<b>Client Sample ID</b>	235784	235785	235786	235787	235788	235789	235790	235791
<b>Sampled Date</b>	27/06/2023 03:07:00 PM	27/06/2023 11:42:00 AM	27/06/2023 02:00:00 PM	27/06/2023 11:10:00 AM	27/06/2023 10:30:00 AM	27/06/2023 09:55:00 AM	27/06/2023 12:15:00 PM	27/06/2023 12:53:00 PM
<b>Sample Number</b>	L23053891	L23053892	L23053893	L23053894	L23053895	L23053896	L23053897	L23053898

**ORGANICS**

TC0049DW : Algal Toxins(Continued)

Microcystin LR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Anatoxin-a(total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Cylindrospermopsin(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Nodularin (total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Date of Performance	DD/MM/YY	30/06/23	30/06/23	30/06/23	30/06/23	30/06/23	30/06/23	30/06/23	30/06/23

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS

<b>Client Sample ID</b>	235784	235785	235786	235787	235788	235789	235790	235791
<b>Sampled Date</b>	27/06/2023 03:07:00 PM	27/06/2023 11:42:00 AM	27/06/2023 02:00:00 PM	27/06/2023 11:10:00 AM	27/06/2023 10:30:00 AM	27/06/2023 09:55:00 AM	27/06/2023 12:15:00 PM	27/06/2023 12:53:00 PM
<b>Sample Number</b>	L23053891	L23053892	L23053893	L23053894	L23053895	L23053896	L23053897	L23053898

**ORGANICS**

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)

Saxitoxin	ug/L	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4
Neosaxitoxin	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
dcSTX	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
C2	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX4	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX3	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
GTX5	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcNeo	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcGTX3	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX6	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX2	ug/L	<1	<1	<1	<1	<1	<1	<1	<1
GTX1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5

<b>Client Sample ID</b>	235784	235785	235786	235787	235788	235789	235790	235791
<b>Sampled Date</b>	27/06/2023 03:07:00 PM	27/06/2023 11:42:00 AM	27/06/2023 02:00:00 PM	27/06/2023 11:10:00 AM	27/06/2023 10:30:00 AM	27/06/2023 09:55:00 AM	27/06/2023 12:15:00 PM	27/06/2023 12:53:00 PM
<b>Sample Number</b>	L23053891	L23053892	L23053893	L23053894	L23053895	L23053896	L23053897	L23053898

**ORGANICS**

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)

C1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcGTX2	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
Date of Performance	DD/MM/YY	03/07/23	03/07/23	03/07/23	03/07/23	03/07/23	03/07/23	03/07/23	03/07/23

<b>Client Sample ID</b>	235800	235801	235802	235803	235804	235805	235806	235807
<b>Sampled Date</b>	27/06/2023 03:07:00 PM	27/06/2023 11:42:00 AM	27/06/2023 02:00:00 PM	27/06/2023 11:10:00 AM	27/06/2023 10:30:00 AM	27/06/2023 09:55:00 AM	27/06/2023 12:15:00 PM	27/06/2023 12:53:00 PM
<b>Sample Number</b>	L23053899	L23053900	L23053901	L23053902	L23053903	L23053904	L23053905	L23053906

**ALGAL**

MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes

Blue Green ASU	ASU/mL	2260	1812	3929	1488	1232	1061	1580	1502
Blue Green Biovol	mm3/L	0.91	1.73	0.933	0.426	0.463	0.305	0.407	0.356
Potentially Toxic Blue Green	cells/mL	0	1040	0	1140	0	0	208	0

<b>Client Sample ID</b>		235800	235801	235802	235803	235804	235805	235806	235807
<b>Sampled Date</b>		27/06/2023 03:07:00 PM	27/06/2023 11:42:00 AM	27/06/2023 02:00:00 PM	27/06/2023 11:10:00 AM	27/06/2023 10:30:00 AM	27/06/2023 09:55:00 AM	27/06/2023 12:15:00 PM	27/06/2023 12:53:00 PM
<b>Sample Number</b>		L23053899	L23053900	L23053901	L23053902	L23053903	L23053904	L23053905	L23053906
<b>ALGAL</b>									
MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes(Continued)									
Potentially Toxic Blue Green ASU	ASU/mL	0.00	60.2	0.00	32.1	0.00	0.00	14.4	0.00
Potentially Toxic Blue Green Biovol	mm3/L	0.000	0.097	0.000	0.031	0.000	0.000	0.024	0.000
Total Blue Green	cells/mL	1167000	867800	2068000	747500	624700	561400	814500	785400
MA91 : Individual Species Total Count, Total BioVol, Total ASU									
Algae Source*	N/A	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL
Date of Performance	DD/MM/YY	13/07/23 00:00	13/07/23 00:00	13/07/23 00:00	13/07/23 00:00	13/07/23 00:00	13/07/23 00:00	13/07/23 00:00	13/07/23 00:00

**COMMENTS**

<u>Sample ID</u>	<u>Comment Level</u>	<u>Method</u>	<u>Test</u>	<u>Comment</u>
L23053899	Method	MA91	-	Debris present in the sample.
L23053900	Method	MA91	-	Debris present in the sample.
L23053901	Method	MA91	-	Debris present in the sample.
L23053902	Method	MA91	-	Debris present in the sample.
L23053903	Method	MA91	-	Debris present in the sample.
L23053904	Method	MA91	-	Debris present in the sample.
L23053905	Method	MA91	-	Debris present in the sample.

\* Indicates NATA accreditation does not cover the performance of this service

---

L23053906	Method	MA91	-	Debris present in the sample.
-----------	--------	------	---	-------------------------------

\* Indicates NATA accreditation does not cover the performance of this service



**LABORATORY QC RESULTS**

N/A - Not Applicable

PQL - Practical Quantitation Limit

LOQ - Limit of Quantification

RPD - Relative Percent Difference

SPIKE/Positive Control - Addition of a known amount and concentration

Duplicate Precision = Accepted - Result 2 within 95% confidence limits of result 1

Duplicate Precision = Outlier - Result 2 outside 95% confidence limits of result 1

Duplicate Precision = Not calculated - Result is outside test range

---

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0049DW Anatoxin-a(extracellular)</b>						
<0.1 ug/L	<0.1	120 50.0 - 120.0 ug/L	120 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0049DW Anatoxin-a(intracellular)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0049DW Anatoxin-a(total)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0049DW Cylindrospermopsin (extracellular)</b>						
<0.05 ug/L	<0.05	93 50.0 - 120.0 ug/L	61 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Cylindrospermopsin (intracellular)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Cylindrospermopsin(total)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0049DW Microcystin LR(extracellular)</b>						
<0.05 ug/L	<0.05	86 50.0 - 120.0 ug/L	70 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin LR(intracellular)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin LR(total)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin RR(extracellular)</b>						
<0.05 ug/L	<0.05	99 50.0 - 120.0 ug/L	74 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin RR(intracellular)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin RR(total)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin YR(extracellular)</b>						
<0.05 ug/L	<0.05	110 50.0 - 120.0 ug/L	82 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0049DW Microcystin YR(intracellular)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin YR(total)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Nodularin (extracellular)</b>						
<0.1 ug/L	<0.1	90 50.0 - 120.0 ug/L	74 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0049DW Nodularin (intracellular)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0049DW Nodularin (total)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0061DW C1</b>						
<0.5 ug/L	<0.5	110 70.0 - 130.0 ug/L	100 % Recovery 50.0 - 130.0 % Recovery	<0.5	<0.5	B 0.0 - 0.0 %
<b>TC0061DW C2</b>						
<0.3 ug/L	<0.3	110 70.0 - 130.0 ug/L	99 % Recovery 50.0 - 130.0 % Recovery	<0.3	<0.3	B 0.0 - 0.0 %

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0061DW dcGTX2</b>						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	110 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
<b>TC0061DW dcGTX3</b>						
<0.3 ug/L	<0.3	110 <i>70.0 - 130.0 ug/L</i>	84 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
<b>TC0061DW dcNeo</b>						
<0.5 ug/L	<0.5	91 <i>70.0 - 130.0 ug/L</i>	70 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
<b>TC0061DW dcSTX</b>						
<0.5 ug/L	<0.5	96 <i>70.0 - 130.0 ug/L</i>	94 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
<b>TC0061DW GTX1</b>						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	91 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
<b>TC0061DW GTX2</b>						
<1 ug/L	<1	110 <i>70.0 - 130.0 ug/L</i>	100 % Recovery <i>50.0 - 130.0 % Recovery</i>	<1	<1	B <i>0.0 - 0.0 %</i>
<b>TC0061DW GTX3</b>						
<0.5 ug/L	<0.5	100 <i>70.0 - 130.0 ug/L</i>	93 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0061DW GTX4</b>						
<0.3 ug/L	<0.3	120 <i>70.0 - 130.0 ug/L</i>	98 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
<b>TC0061DW GTX5</b>						
<0.5 ug/L	<0.5	97 <i>70.0 - 130.0 ug/L</i>	98 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
<b>TC0061DW GTX6</b>						
<0.3 ug/L	<0.3	110 <i>70.0 - 130.0 ug/L</i>	110 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
<b>TC0061DW Neosaxitoxin</b>						
<0.3 ug/L	<0.3	98 <i>70.0 - 130.0 ug/L</i>	73 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
<b>TC0061DW Saxitoxin</b>						
<0.4 ug/L	<0.4	97 <i>70.0 - 130.0 ug/L</i>	92 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.4	<0.4	B <i>0.0 - 0.0 %</i>

\* Indicates NATA accreditation does not cover the performance of this service

## Extra Note:

F: Blank is not applicable for this analyte

E: Spike is not applicable for this analyte

DUPLICATE Anatoxin-a(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcNeo B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcSTX B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX4 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX5 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX6 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Neosaxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Saxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

\* Indicates NATA accreditation does not cover the performance of this service

\* Indicates NATA accreditation does not cover the performance of this service